



# UNITED STATES PATENT AND TRADEMARK OFFICE

UNITED STATES DEPARTMENT OF COMMERCE  
United States Patent and Trademark Office  
Address: COMMISSIONER FOR PATENTS  
P.O. Box 1450  
Alexandria, Virginia 22313-1450  
www.uspto.gov

APPLICATION NO.	FILING DATE	FIRST NAMED INVENTOR	ATTORNEY DOCKET NO.	CONFIRMATION NO.
10/752,933	01/07/2004	Peter M. Bonutti	2500DV2CN2DV3CN8	5806
7590 12/22/2008				
Patent Counsel United States Surgical, a division of TYCO HEALTHCARE GROUP LP 150 Glover Avenue Norwalk, CT 06856		EXAMINER TRUONG, KEVIN THAO		
		ART UNIT PAPER NUMBER 3734		
		MAIL DATE DELIVERY MODE 12/22/2008 PAPER		

**Please find below and/or attached an Office communication concerning this application or proceeding.**

The time period for reply, if any, is set in the attached communication.

### Office Action Summary

**Application No.**

10/752,933

**Applicant(s)**

BONUTTI, PETER M.

**Examiner**

Kevin T. Truong

**Art Unit**

3734

-- The MAILING DATE of this communication appears on the cover sheet with the correspondence address --  
**Period for Reply**

A SHORTENED STATUTORY PERIOD FOR REPLY IS SET TO EXPIRE 3 MONTH(S) OR THIRTY (30) DAYS, WHICHEVER IS LONGER, FROM THE MAILING DATE OF THIS COMMUNICATION.

- Extensions of time may be available under the provisions of 37 CFR 1.136(a). In no event, however, may a reply be timely filed after SIX (6) MONTHS from the mailing date of this communication.
- If NO period for reply is specified above, the maximum statutory period will apply and will expire SIX (6) MONTHS from the mailing date of this communication.
- Failure to reply within the set or extended period for reply will, by statute, cause the application to become ABANDONED (35 U.S.C. § 133). Any reply received by the Office later than three months after the mailing date of this communication, even if timely filed, may reduce any earned patent term adjustment. See 37 CFR 1.704(b).

**Status**

- 1) ☒ Responsive to communication(s) filed on RCE and Amendt. 10/15/2008.  
2a) ☒ This action is **FINAL**. 2b) ☐ This action is non-final.  
3) ☐ Since this application is in condition for allowance except for formal matters, prosecution as to the merits is closed in accordance with the practice under *Ex parte Quayle*, 1935 C.D. 11, 453 O.G. 213.

**Disposition of Claims**

- 4) ☒ Claim(s) 1 and 4-18 is/are pending in the application.  
4a) Of the above claim(s) \_\_\_\_\_ is/are withdrawn from consideration.  
5) ☐ Claim(s) \_\_\_\_\_ is/are allowed.  
6) ☒ Claim(s) 1 and 4-18 is/are rejected.  
7) ☐ Claim(s) \_\_\_\_\_ is/are objected to.  
8) ☐ Claim(s) \_\_\_\_\_ are subject to restriction and/or election requirement.

**Application Papers**

- 9) ☐ The specification is objected to by the Examiner.  
10) ☐ The drawing(s) filed on \_\_\_\_\_ is/are: a) ☐ accepted or b) ☐ objected to by the Examiner.  
Applicant may not request that any objection to the drawing(s) be held in abeyance. See 37 CFR 1.85(a).  
Replacement drawing sheet(s) including the correction is required if the drawing(s) is objected to. See 37 CFR 1.121(d).  
11) ☐ The oath or declaration is objected to by the Examiner. Note the attached Office Action or form PTO-152.

**Priority under 35 U.S.C. § 119**

- 12) ☐ Acknowledgment is made of a claim for foreign priority under 35 U.S.C. § 119(a)-(d) or (f).  
a) ☐ All b) ☐ Some \* c) ☐ None of:  
1. ☐ Certified copies of the priority documents have been received.  
2. ☐ Certified copies of the priority documents have been received in Application No. \_\_\_\_\_.  
3. ☐ Copies of the certified copies of the priority documents have been received in this National Stage application from the International Bureau (PCT Rule 17.2(a)).

\* See the attached detailed Office action for a list of the certified copies not received.

**Attachment(s)**

- 1) ☐ Notice of References Cited (PTO-892)  
2) ☐ Notice of Draftsperson's Patent Drawing Review (PTO-948)  
3) ☐ Information Disclosure Statement(s) (PTO/S508)  
Paper No(s)/Mail Date \_\_\_\_\_  
4) ☐ Interview Summary (PTO-413)  
Paper No(s)/Mail Date \_\_\_\_\_  
5) ☐ Notice of Informal Patent Application  
6) ☐ Other: \_\_\_\_\_

## **DETAILED ACTION**

### ***Continued Examination Under 37 CFR 1.114***

1. A request for continued examination under 37 CFR 1.114, including the fee set forth in 37 CFR 1.17(e), was filed in this application after final rejection. Since this application is eligible for continued examination under 37 CFR 1.114, and the fee set forth in 37 CFR 1.17(e) has been timely paid, the finality of the previous Office action has been withdrawn pursuant to 37 CFR 1.114. Applicant's submission filed on 10/15/2008 has been entered.

### ***Claim Rejections - 35 USC § 103***

1. The following is a quotation of 35 U.S.C. 103(a) which forms the basis for all obviousness rejections set forth in this Office action:

(a) A patent may not be obtained though the invention is not identically disclosed or described as set forth in section 102 of this title, if the differences between the subject matter sought to be patented and the prior art are such that the subject matter as a whole would have been obvious at the time the invention was made to a person having ordinary skill in the art to which said subject matter pertains. Patentability shall not be negated by the manner in which the invention was made.

2. Claims 1 and 4-18 are rejected under 35 U.S.C. 103(a) as being unpatentable over Matthews et al. (US 3253594) in view of Debbas (US 4966583).

As to claims 1, 4-9, and 11-18, Matthews et al. disclose a cannula (the trocar described in col. 4, lines 39-47) and a retractor including shaft 16 that has a first lumen 56 therethrough and inflatable bladder 18 attached to a distal end of the shaft 16 and extending distally thereof, the inflatable bladder 18 including an opening extending therethrough (the opening through which instrument 12 passes) such that a surgical instrument 12 may traverse the first lumen 56 and the opening thereby entering the working space. The bladder 18 is inherently capable of retracting opposed surfaces of a

Art Unit: 3734

bone joint since the inflation of the bladder would inherently apply an outward pressure against the opposed surfaces to retract them. Alternatively, it would have been obvious that inflatable bladder 18 is capable of retracting opposed surfaces of a bone joint for the reasons set forth above. As to claim 4, bladder 43 is eccentrically shaped or mounted on the shaft as indicated in col. 3, lines 9-19. As to claim 5, note col. 2, lines 41-43. As to claim 8, portion 38 may be considered to be part of the shaft. This portion 38 of the shaft is rigid.

Matthews et al does not disclose the inflatable bladder is inflated by fluid received through a second lumen defined between the inner and outer wall of the shaft.

However, Debbas teaches that it is known in the surgical art to have inflatable bladder is inflated by fluid received through a second lumen defined between the inner and outer wall of the shaft.

It would have been obvious to one of ordinary skill in the art at the time the invention was made to make the shaft of Matthews et al with inner and outer walls so that the inflatable bladder can be inflated by fluid received through a lumen defined between the inner and outer walls of the shaft as taught by Debbas for cost effective. Claim 10 is rejected under 35 U.S.C. 103(a) as being unpatentable over Matthews et al. (3,253,594). Matthews et al. fail to disclose a viewing scope. However, Matthews et al. teach that member 12 can by any one of a number of devices, one of which can be used for taking photographs of interior portions of the peritoneal cavity (col. 2, lines 49-53). It is old and well known in the art that a viewing scope should be part of a device for taking photographs of interior portions of the body apparently in order to obtain the

Art Unit: 3734

advantage of viewing the tissue being photographed. It would have been obvious to use a viewing scope as the Matthews et al. member 12 in the so that it too would have this advantage.

### ***Response to Arguments***

2. Applicant's arguments filed 10/15/2008 have been fully considered but they are not persuasive. In response to applicant's argument that there is no suggestion to combine the references, the examiner recognizes that obviousness can only be established by combining or modifying the teachings of the prior art to produce the claimed invention where there is some teaching, suggestion, or motivation to do so found either in the references themselves or in the knowledge generally available to one of ordinary skill in the art. See *In re Fine*, 837 F.2d 1071, 5 USPQ2d 1596 (Fed. Cir. 1988) and *In re Jones*, 958 F.2d 347, 21 USPQ2d 1941 (Fed. Cir. 1992). In this case, modifying Matthews et al device to inflate the balloon through a second lumen defined between the inner and outer wall of the shaft as taught by Debbas patent, rather than the exterior lumen would have been an obvious to one of ordinary skill in the art at the time the invention was made to so. Although applicant may recognize another advantage which would flow naturally from following the suggestion of the prior art cannot be the basis for patentability when the differences would otherwise be obvious. See *Ex parte Obiaya*, 227 USPQ 58, 60 (Bd. Pat. App. & Inter. 1985).

### ***Conclusion***

This is a RCE of applicant's earlier Application No. 10/752,933. All claims are drawn to the same invention claimed in the earlier application and could have been

Art Unit: 3734

finally rejected on the grounds and art of record in the next Office action if they had been entered in the earlier application. Accordingly, **THIS ACTION IS MADE FINAL** even though it is a first action in this case. See MPEP § 706.07(b). Applicant is reminded of the extension of time policy as set forth in 37 CFR 1.136(a).

A shortened statutory period for reply to this final action is set to expire **THREE MONTHS** from the mailing date of this action. In the event a first reply is filed within **TWO MONTHS** of the mailing date of this final action and the advisory action is not mailed until after the end of the **THREE-MONTH** shortened statutory period, then the shortened statutory period will expire on the date the advisory action is mailed, and any extension fee pursuant to 37 CFR 1.136(a) will be calculated from the mailing date of the advisory action. In no, however, event will the statutory period for reply expire later than **SIX MONTHS** from the mailing date of this final action.

3. Any inquiry concerning this communication or earlier communications from the examiner should be directed to Kevin T. Truong whose telephone number is 571-272-4705. The examiner can normally be reached on Monday-Friday from 8:00 AM to 4:30 PM..

The fax phone number for the organization where this application or proceeding is assigned is 571-273-8300.

Art Unit: 3734

Information regarding the status of an application may be obtained from the Patent Application Information Retrieval (PAIR) system. Status information for published applications may be obtained from either Private PAIR or Public PAIR. Status information for unpublished applications is available through Private PAIR only. For more information about the PAIR system, see <http://pair-direct.uspto.gov>. Should you have questions on access to the Private PAIR system, contact the Electronic Business Center (EBC) at 866-217-9197 (toll-free). If you would like assistance from a USPTO Customer Service Representative or access to the automated information system, call 800-786-9199 (IN USA OR CANADA) or 571-272-1000.

/Kevin T. Truong/  
Primary Examiner, Art Unit 3734

Kevin T. Truong  
Primary Examiner  
Art Unit 3734